



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/789,610	02/27/2004	Roy K. Lim	4002-3438	9038

7590 05/16/2006

Woodard, Emhardt, Moriarty, McNett & Henry LLP  
Bank One Center/Tower  
Suite 3700  
111 Monument Circle  
Indianapolis, IN 46204-5137

EXAMINER

KIM, JOHN

ART UNIT PAPER NUMBER

3733

DATE MAILED: 05/16/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

<b>Office Action Summary</b>	Application No. 10/789,610	Applicant(s) LIM, ROY K.	
	Examiner John Kim	Art Unit 3733	

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☐ Responsive to communication(s) filed on \_\_\_\_.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-31 and 38-68 is/are pending in the application.  
     4a) Of the above claim(s) \_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1, 2 and 38-41 is/are rejected.
- 7) ☒ Claim(s) 3-31 and 42-68 is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 27 February 2004 is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
     Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
     Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  
     a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)  | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. ____. |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                   | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)             |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date ____. | 6) <input type="checkbox"/> Other: ____.  |

## **DETAILED ACTION**

### ***Drawings***

The drawings are objected to under 37 CFR 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the "sleeve at least partially surrounding said rod" as in claim 1 and "sleeve includes an end portion adapted to contact at least one of the group consisting of a spinal rod, an orthopedic connector, and an orthopedic plate" as in claim 40 must be shown or the feature(s) canceled from the claim(s). No new matter should be entered.

Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. The figure or figure number of an amended drawing should not be labeled as "amended." If a drawing figure is to be canceled, the appropriate figure must be removed from the replacement sheet, and where necessary, the remaining figures must be renumbered and appropriate changes made to the brief description of the several views of the drawings for consistency. Additional replacement sheets may be necessary to show the renumbering of the remaining figures. Each drawing sheet submitted after the filing date of an application must be labeled in the top margin as either "Replacement Sheet" or "New Sheet" pursuant to 37 CFR 1.121(d). If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

***Claim Rejections - 35 USC § 102***


The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1 and 2 are rejected under 35 U.S.C. 102(b) as being anticipated by Krafft (US Pat 2681807).

In regards to claim 1, Krafft discloses an apparatus having a base (26), a shaft (4) that is connected to the base (by 38, 34, and 28) and being slidable with respect to the shaft, a plate (18) with an aperture (which allows the fuse to enter) where the plate has two positions (figure 1 and 2) where figure 2 shows the plate in the "locked" position, one of which locks the implant in place, and a sleeve (21, 23) which partially surrounds the sleeve and is slidable with respect to the base and plate. In regards to claim 2, Krafft discloses having a ratchet mechanism (17), where the coils of the spring are the teeth of a bar and a pin (A, see figure below) is the pawl. The ratchet is connected to the base and sleeve, and is operable to move the sleeve by pushing the plate (18) which is connected to the sleeve at (24).



Claims 38-41 are rejected under 35 U.S.C. 102(b) as being anticipated by Krafft (US Pat 2681807) with a different interpretation.

In regards to claim 38, Krafft has a plate (18) with two edges and an aperture, a shaft (21, 23) which is connected to the plate at one edge (near 16) and can pivot the plate around the second edge (near 18). The plate moves between two positions (figures 1 and 2) where figure 2 shows the plate in the "locked" position. In regards to claims 39 and 40, Krafft discloses having a sleeve (7), which is slidable with respect to the shaft, and the sleeve has an end portion which is able to contact a spinal rod. As discussed above and in reference to claim 41, Krafft discloses having a ratchet mechanism (17), where the coils of the spring are the teeth of a bar and a pin (A, see figure above) is the pawl.

#### ***Allowable Subject Matter***

Claims 3-31 and 42-68 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

***Response to Arguments***

Applicant's arguments with respect to claims 1-31 have been considered but are moot in view of the new ground(s) of rejection.

***Conclusion***

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. See PTO-892 for art cited of interest.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John Kim whose telephone number is (571) 272-2817. The examiner can normally be reached on M-F 8-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

JK 

  
EDUARDO C. ROBERT  
SUPERVISORY PATENT EXAMINER